

EU Declaration of Conformity

Application of Council EMC Directive 89/336/EEC.

NOTE

The Low Voltage Directive 73/23/EEC does not apply to the Model 3430 Plus and 3440 Plus Surface Moisture-Density Gauges because there are no voltages greater than 75 V dc that are accessible by the operator.

Standards to Which Conformity is Declared:

Guidance Documents:

Emissions: EN 61326-2:1997 +A1:1998 +A2:2001
Immunity: EN 61326-2:1997 +A1:1998 +A2:2001

Test Methods:

Emissions: EN 55011:1998, EN 61000-3-2:2000, EN 61000-3-3:1995 +A1:2001
Immunity: EN 61000-4-2:1995 +A1:1998 +A2:2001, EN 61000-4-3:2002, EN 61000-4-4:1995, EN 61000-4-5:1995 +A1:1996, EN 61000-4-6:1996 +A1:2001, EN 61000-4-11:2001

Manufacturer: Troxler Electronic Laboratories, Inc.
P.O. Box 12057
3008 Cornwallis Road
Research Triangle Park, North Carolina
27709
U.S.A.

Apparatus: Model 3430 Plus and 3440 Plus
Surface Moisture-Density Gauges


Type of Equipment: Measurement, Control, and Laboratory Use

Class of Equipment: Class A

Year of Declaration: 2006

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Signature:



Date:

19 March 2007

Full Name: William F. Troxler, Jr.

Title: President & Chairman of the Board

Place of issue: Research Triangle Park, NC, U.S.A.